



**NATIONAL INSTITUTE OF TECHNOLOGY  
ROURKELA-769008 (ODISHA)**  
(An Institute of National Importance under Ministry of HRD, GOI)

**NOTICE INVITING TENDER**

**Tender Notification No: NITR/PW/EE/2018/60**

**Dated: 26/09/2018**

The National Institute of Technology, Rourkela invites bids from the eligible bidders for providing the service for **Calibration of High Voltage Laboratory Equipments for NABL Accreditation.**

**Last date of Submission of Bid : 25.10.2018 at 11:00 AM**

**Date of opening of Technical and Financial Bid : 26.10.2018 at 11:00 AM**

**For Details: [http://nitrkl.ac.in/OldWebsite/Jobs\\_Tenders/8Maintenance/Default.aspx](http://nitrkl.ac.in/OldWebsite/Jobs_Tenders/8Maintenance/Default.aspx)**

**Contact: Dr. Subrata Karmakar, EE; Ph. 0661-2462411;**

**Email: [karmakars@nitrkl.ac.in](mailto:karmakars@nitrkl.ac.in)**

**Bidding through: <https://eprocure.gov.in/eprocure/app>**

**sd  
REGISTRAR**



# NATIONAL INSTITUTE OF TECHNOLOGY ROURKELA – 769 008, ODISHA

**(OPEN TENDER NOTICE NO: NITR/PW/EE/2018/60 Date: 26.09.2018)**  
**Calibration of High Voltage Laboratory Equipments for NABL Accreditation**

SI.No	Description of Goods/Service	Quantity																
1.	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th colspan="4" style="text-align: left;">500kV AC Power Frequency Test Setup.</th> </tr> <tr> <th style="width: 10%;">SL No.</th> <th style="width: 40%;">Name of Equipment</th> <th style="width: 20%;">Model/Type</th> <th style="width: 30%;">Range/Ratings</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">1</td> <td>Power Frequency High Voltage Testing Transformer cascaded of two 250kVA transformer with Coupling Capacitor and Measuring Capacitor and a Sphere Gap.</td> <td>Make: WSTS Test Systems, Type-TEOK 500-500</td> <td>0.6/500kV, 500 kVA, 2 Phase, 50 Hz PD Level &lt;5pC</td> </tr> </tbody> </table>	500kV AC Power Frequency Test Setup.				SL No.	Name of Equipment	Model/Type	Range/Ratings	1	Power Frequency High Voltage Testing Transformer cascaded of two 250kVA transformer with Coupling Capacitor and Measuring Capacitor and a Sphere Gap.	Make: WSTS Test Systems, Type-TEOK 500-500	0.6/500kV, 500 kVA, 2 Phase, 50 Hz PD Level <5pC	1.00				
500kV AC Power Frequency Test Setup.																		
SL No.	Name of Equipment	Model/Type	Range/Ratings															
1	Power Frequency High Voltage Testing Transformer cascaded of two 250kVA transformer with Coupling Capacitor and Measuring Capacitor and a Sphere Gap.	Make: WSTS Test Systems, Type-TEOK 500-500	0.6/500kV, 500 kVA, 2 Phase, 50 Hz PD Level <5pC															
2.	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th colspan="4" style="text-align: left;">400kV Impulse Voltage Test Setup.</th> </tr> <tr> <th style="width: 10%;">SL No.</th> <th style="width: 20%;">Name of Equipment</th> <th style="width: 20%;">Model/Type</th> <th style="width: 50%;">Range/Ratings</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">1</td> <td>Impulse Voltage Generator</td> <td>Make:Haefly, USA Type:33</td> <td>400kV, 20 kJ, 2 Stage with output of either +ve or -ve polarity Output wave shape 1.2 / 50 <math>\mu</math>s with permissible tolerance as per IS 2071 and Output wave shape 250 / 2500 <math>\mu</math>s with permissible tolerance as per IS 2071</td> </tr> <tr> <td style="text-align: center;">1.a</td> <td>Peak Voltmeter</td> <td>Make: Haefly, USA Type-65</td> <td>200/400/800/1600 V</td> </tr> </tbody> </table>	400kV Impulse Voltage Test Setup.				SL No.	Name of Equipment	Model/Type	Range/Ratings	1	Impulse Voltage Generator	Make:Haefly, USA Type:33	400kV, 20 kJ, 2 Stage with output of either +ve or -ve polarity Output wave shape 1.2 / 50 $\mu$ s with permissible tolerance as per IS 2071 and Output wave shape 250 / 2500 $\mu$ s with permissible tolerance as per IS 2071	1.a	Peak Voltmeter	Make: Haefly, USA Type-65	200/400/800/1600 V	1.00
400kV Impulse Voltage Test Setup.																		
SL No.	Name of Equipment	Model/Type	Range/Ratings															
1	Impulse Voltage Generator	Make:Haefly, USA Type:33	400kV, 20 kJ, 2 Stage with output of either +ve or -ve polarity Output wave shape 1.2 / 50 $\mu$ s with permissible tolerance as per IS 2071 and Output wave shape 250 / 2500 $\mu$ s with permissible tolerance as per IS 2071															
1.a	Peak Voltmeter	Make: Haefly, USA Type-65	200/400/800/1600 V															
3.	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th colspan="4" style="text-align: left;">100kV AC power frequency test Setup for partial discharge (PD) measurement.</th> </tr> <tr> <th style="width: 10%;">SL No.</th> <th style="width: 40%;">Name of Equipment</th> <th style="width: 20%;">Model/Type</th> <th style="width: 30%;">Range/Ratings</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">1</td> <td>Power Frequency High Voltage Testing Transformer with Coupling Capacitor, Measuring Capacitor, PD Meter, PD pulse generator and doubled shelled room.</td> <td>Make: WSTS Test Systems, Type-TEO 100/20</td> <td>0.22/100kV, 10 kVA, 0.5 A, 1 Phase, 50 Hz PD Level &lt;5pC</td> </tr> </tbody> </table>	100kV AC power frequency test Setup for partial discharge (PD) measurement.				SL No.	Name of Equipment	Model/Type	Range/Ratings	1	Power Frequency High Voltage Testing Transformer with Coupling Capacitor, Measuring Capacitor, PD Meter, PD pulse generator and doubled shelled room.	Make: WSTS Test Systems, Type-TEO 100/20	0.22/100kV, 10 kVA, 0.5 A, 1 Phase, 50 Hz PD Level <5pC	1.00				
100kV AC power frequency test Setup for partial discharge (PD) measurement.																		
SL No.	Name of Equipment	Model/Type	Range/Ratings															
1	Power Frequency High Voltage Testing Transformer with Coupling Capacitor, Measuring Capacitor, PD Meter, PD pulse generator and doubled shelled room.	Make: WSTS Test Systems, Type-TEO 100/20	0.22/100kV, 10 kVA, 0.5 A, 1 Phase, 50 Hz PD Level <5pC															

**Note: 1) Organization/Bidder Should be NABL Certified Laboratory 2) Price including fooding/ Lodging /installation etc. 3) Aforesaid three items will include all prices.**

**Quantity required Delivery:** As mentioned above (All information provided in technical specification)

**Delivery:** Within 60 days from the date of purchase order

**Last Date of submission of Tender** : 25.10.2018 at 11: 00 AM

**Date of opening of Technical and Financial Bid** : 26.10.2018 at 11:00 AM

1. The firm should not have been black listed at any time.
2. The submission of following bids by the tenderer should be through <https://eprocure.gov.in/eprocure/app>
3. Please follow the guidelines as per the portal.

**Calibration of High Voltage Laboratory Equipments for NABL Accreditation  
(Tender Notice No.: - NITR/PW/EE/2018/ , dated: 26.09.2018 Due on  
25.10.2018 at 11:00 AM )**

4. Liquidated damage clause will be charged for any delay in supply of goods.
5. The validity of the tender shall be **90 days** from the date of opening of the bids.
6. Detailed advertisement including all tender documents is also available in our website at [http://nitrkl.ac.in/OldWebsite/Jobs\\_Tenders/9Equipment/Default.aspx](http://nitrkl.ac.in/OldWebsite/Jobs_Tenders/9Equipment/Default.aspx).
7. NIT reserves the right to qualify or deny prequalification of any or all applicants without assigning any reasons.

**(REGISTRAR)**  
**NIT, Rourkela**  
**Fax No- 0661-2462022**  
**Ph. No -0661-2472021**